

C57023



GEI INTERNATIONAL, INC.
100 Ball Street, P.O. Box 6849, Syracuse, NY 13217-6849 USA
Telephone: (315) 463-9261 Fax: (315) 463-9034
Website: www.geionline.com

Certificate of Calibration

Stainless Steel Ruler Model 2020A, 6" Length

Serial Number: C57023
Date of Certification: 09/10/15
Temperature (Fahrenheit): 71 Degrees
Relative Humidity (RH): 68%

GEI International, Inc., as the OEM manufacturer certifies that this stainless steel ruler with the serial number as indicated above is accurate within the tolerances as noted below on the date of certification.

Inch (") and Millimeter (mm) Scales:

SAMPLE

From the center of the first etched graduation to the center of the last etched graduation, the scale error is within plus or minus 0.005" of each total scale length.

The equipment used for the final inspection of this stainless steel ruler is certified traceable to master standards at the National Institute of Standards and Technology (NIST), Gaithersburg, MD.

The test method used was the direct measurement of the lengths as described above using Certified Master Vernier Calipers certified to specification GGG-C-111C including National Institute of Standards and Technology Test #683/282266. The serial number of the Certified Master Vernier Calipers used is 955713.

Inspection Manager: _____

A handwritten signature in black ink, appearing to read 'J. Paul', is written over a horizontal line that serves as a signature line.