



GEI INTERNATIONAL, INC.

100 Ball Street, P.O. Box 6849, Syracuse, NY 13217-6849 USA

Telephone: (315) 463-9261 Fax: (315) 463-9034

Website: www.geionline.com

Certificate of Calibration

Stainless Steel Ruler Model 2020A, 6" Length

Date of Manufacture: 03/12/10

Serial Number: C38010

We certify that the above stainless steel ruler with the serial number as indicated is accurate within the tolerances noted below on the date of manufacture.

Inch (") and Millimeter (mm) Scales:

SAMPLE

From the center of the first etched graduation to the center of the last etched graduation, the scale error is within plus or minus 0.005" of each scale length.

The equipment used for the final inspection of this stainless steel ruler is certified traceable to master standards at the National Institute of Standards and Technology (NIST), Washington, DC.

The test method used was the direct measurement of the lengths as described above using Certified Master Vernier Calipers certified to specification GGG-C-111C including National Institute of Standards and Technology Test # 821/273187 The serial number of the Certified Master Vernier Calipers used is 10052546.

Inspection Manager: _____

A handwritten signature in black ink, appearing to read "J. Paul", is written over a horizontal line.

Manufacturers of Stainless Steel Rulers & Gauges

www.geionline.com